

Code No.29-5180

# SPECIFICATION

## Limulus Amebocyte Lysate ES-III, lyophilized for Endotoxin Detection

REQUIREMENT	SPECIFICATION
Appearance	to pass test
Solubility in water	to pass test
Sensitivity	0.015~0.06EU/mL
Performance test (Keinetic turbidimetric analysis)	to pass test
Stability of solution	to pass test

Manufacturer FUJIFILM Healthcare Manufacturing Corporation